



AC200031
AGENCY DOCUMENTS, ET1928L (3rd Edition)

REVISION HISTORY

Rev	Description	Author.	Approved by	Date
A	Initial release per ECO-12-016856	Dean King	Dean King	September 28, 2012

DOC #:
AC200031
Print Date: 28-Sep-12

REV:
A

Page 1 of
26

**UNCONTROLLED COPY
OF RELEASED DOCUMENT FROM
DOCUMENTATION SERVICES**



Contents

Elo Declaration of Conformity (1 page).

ITE CB Certificate - Intertek (1 page).

Medical CB Certificate – TUV (1 page).

UL Cover Sheet - ANSI/AAMI ES60601-1 (2005 + C1:09 + A2:10) and CAN/CSA-C22.2 No. 60601-1 (2008) (3 pages).

UL Cover Sheet - UL 60601-1, 1st Edition, 2006-04-26 and CAN/CSA-C22.2 No. 601.1-M90, 2005 (4 pages) (4 pages).

CCC Certificate (6 pages).

EMC certificate – Information Technology (1 page).

EMC certificate – Medical (1 page).

FCC and ICES certificate (1 page).

KC certificate (1 page).

BSMI (2 pages).

VCCI (2 pages).

Declaration of Conformity


Application of Council Directives: 2004/108/EC, 2006/95/EC, & 93/68/EEC

Standards to which conformity is declared:

IEC 60950-1:2001, 2nd Ed. and/or EN 60950-1:2006+A11
IEC 60601-1:2005 (3rd Edition) + CORR. 1 (2006) + CORR. 2 (2007)
ANSI/AAMI ES 60601-1: 2005, First Edition
CAN/CSA-C22.2 No. 60601-1 (2008)
UL 60601-1, 1st Edition, 2006-04-26
CAN/CSA-C22.2 No. 601.1-M90, 2005
CISPR 22: 2008 Class B
EN 55022: 2010 Class B ITE
AS/NZS CISPR22:2009 Class B ITE
CISPR24:2010
EN 55024: 2010
FCC Part 15: Subpart B, Class B
ICES-003: 2004, Class B
V-3/2007.04, Class B
EN 61000-6-1:2007
EN 60601-1-2: 2007
CISPR 11:2009 +A1:2010 (Group I, Class B)
AS/NZS CISPR 11:2011 (Group I, Class B)
EN 55011:2009 +A1:2010, (Group I, Class B)
IEC 61000-3-2:2005 +A1:2008 +A2:2009 (The power consumption is less than 75W and no limits apply.)
IEC 61000-3-3:2008
IEC 61000-4-2: 2008 ED.2.0, 6 kV contact, 8 kV air
IEC 61000-4-3: 2010 ED.3.2, 80-2500 MHz, 10 V/m, 80% AM (1 KHz)
IEC 61000-4-4: 2012 ED.3.0, 2,0/1,0 kV (power/IO) at 5 KHz
IEC 61000-4-5: 2005 ED.2.0, 1,0/2,0 kV L-L, L-E
IEC 61000-4-6: 2008 ED.3.0, 0,15 – 80 MHz, 3 Vrms, 80% AM, 1 KHz
IEC 61000-4-8: 2009 ED.2.0, 3 A/m 50-60 Hz
IEC 61000-4-11: 2004 ED.2.0, 0/40/70% U_T, dips, variations
ENV 50204: 1995, 900 MHz, 3 V/m
U.S. Rehabilitation Act Section 508

Manufacturer's Name and Address:	Elo Touch Solutions, Inc. 301 Constitution Drive Menlo Park, California, USA 94025-1164
Manufacturer's Address in EU:	Elo Touch Solutions (Belgium) NV Diestsesteenweg 692 B-3010 Kessel-Lo Belgium
Type of Equipment	LCD Touch Monitor
Product Name:	48,3 cm (19 in.) LCD Touch Monitor
Model Number(s)	ET1928L-XYZM* (X=0-9/A-Z, Y= C/N/S/U, Z=A-Z, *= suffixes)
First Year of Manufacture :	2012

I hereby declare that the equipment specified above conforms to the above Directives and Standard(s).



Craig Schwarz
Director of Engineering

9/28/2012



Ref. Certif. No.

SE-62506M2

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST
CERTIFICATES FOR ELECTRICAL EQUIPMENT
(IECEE) CB SCHEME

SYSTEME CEI D'ACCEPTATION MUTUELLE DE
CERTIFICATS D'ESSAIS DES EQUIPEMENTS
ELECTRIQUES (IECEE) METHODE OC

CB TEST CERTIFICATE

CERTIFICAT D'ESSAI OC

Product
Produit

LCD Touch Monitor

Name and address of the applicant
Nom et adresse du demandeur

Elo TouchSystems, Inc.
301 Constitution Drive, Menlo Park
CA 94025, U.S.A.

Name and address of the manufacturer
Nom et adresse du fabricant

Same as applicant

Name and address of the factory
Nom et adresse de l'usine
Note: When more than one factory, please report on page 2
Note: Lorsque il y plus d'une usine, veuillez utiliser la 2^{ème} page

See page 2

Ratings and principal characteristics
Valeurs nominales et caractéristiques principales

12 Vdc, 4 A
Class III

Trademark (if any)
Marque de fabrique (si elle existe)

Elo

Type of Manufacturer's Testing Laboratories used
Type de programme du laboratoire d'essais
constructeur

Model / Type Ref.
Ref. De type

ET1928L-***A-*, ET1928L-***M-*,
ET1928L-***N-*

Additional information (if necessary may also be
reported on page 2)
Les informations complémentaires (si nécessaire,
peuvent être indiqués sur la 2^{ème} page)

See page 2

A sample of the product was tested and found
to be in conformity with
Un échantillon de ce produit a été essayé et a été
considéré conforme à la

IEC 60950-1:2005. Group and national differences for the
CENELEC countries (EN 60950-1:2006 + A11)

As shown in the Test Report Ref. No. which forms part
of this Certificate
Comme indiqué dans le Rapport d'essais numéro de
référence qui constitue partie de ce Certificat

TP10070141-ST, TP10070141-ST(R1) and TP10070141-ST(R2)

This CB Test Certificate is issued by the National Certification Body
Ce Certificat d'essai OC est établi par l'Organisme National de Certification

Intertek Semko AB
Box 1103
SE-164 22 Kista, Sweden
Int +46 8 750 00 00



Signature:

Bo Berglöv

Date: 15 March 2012

IEC SYSTEM FOR MUTUAL RECOGNITION OF TEST
CERTIFICATES FOR ELECTRICAL EQUIPMENT
(IECEE) CB SCHEME

SYSTEME CEI D'ACCEPTATION MUTUELLE DE
CERTIFICATS D'ESSAIS DES EQUIPEMENTS
ELECTRIQUES (IECEE) METHODE OC

CB TEST CERTIFICATE
CERTIFICAT D'ESSAI OC

Product
Produit

LCD Monitor

Name and address of the applicant
Nom et adresse du demandeur

Elo TouchSystems, Inc.
301 Constitution Drive
Menlo Park CA 94025-1111, USA

Name and address of the manufacturer
Nom et adresse du fabricant

Elo TouchSystems, Inc.
301 Constitution Drive
Menlo Park CA 94025-1111, USA

Name and address of the factory
Nom et adresse de l'usine

☒ Additional Information page 2

Note: When more than one factory, please report on page 2
Note: Lorsque il y plus d'une usine, veuillez utiliser la 2^{ème} page

Ratings and principal characteristics
Valeurs nominales et caractéristiques principales

DC 12V; 4A

Trade mark (if any)
Marque de fabrique (si elle existe)

TE connectivity

Model/type Ref.
Ref. de type

ET1928L-XXXN-X-XXXXXXXXXXXXXXXXXX
(X = A-Z, 0-9, hyphen or blank for marketing purpose)
ET1928L-XXXM-X-XXXXXXXXXXXXXXXXXX
(X = A-Z, 0-9, hyphen or blank for marketing purpose)

Additional information (if necessary may also be
reported on page 2)
Les Information complémentaire (si nécessaire,
peuvent être indiqués sur la 2^{ème} page)

Models differ in model name only.

A sample of the product was tested and found
to be in conformity with
Un échantillon de ce produit a été essayé et a été
considéré conforme à la

PUBLICATION **EDITION**

IEC 60601-1:2005
for national deviations see test report

As shown in the Test Report Ref. No. which forms part
of this Certificate
Comme indiqué dans le Rapport d'essais numéro de
référence qui constitue une partie de ce Certificat

11028194 001

This CB Test Certificate is issued by the National Certification Body
Ce Certificat d'essai OC est établi par l'Organisme National de Certification



TÜV Rheinland LGA Products GmbH
Tillystraße 2 · 90431 Nürnberg, Germany
Phone + 49 221 806-1371
Fax + 49 221 806-3935
Mail: cert-validity@de.tuv.com
Web: www.tuv.com



Date: 30.04.2012

Signature:

Dipl.-Ing. W. Hsu

UL TEST REPORT AND PROCEDURE

Standard:	ANSI/AAMI ES60601-1 (2005 + C1:09 + A2:10)(Medical Electrical Equipment - Part 1: General Requirements for Basic Safety and Essential Performance) CAN/CSA-C22.2 No. 60601-1 (2008) (Medical Electrical Equipment - Part 1: General Requirements for Basic Safety and Essential Performance)
Certification Type:	Classification
CCN:	MDAF, MDAF7 (Medical Equipment - General Medical Equipment)
Product:	LCD Desktop Touchmonitor
Model:	ET1928L-XXXM-XXXXXXXXXXXXXXXXXXXXX (X Can be 0-9, A-Z or blank, X for marking purpose, not relates to any safety issue)
Rating:	Adaptor Input 100-240 V ac, 60/50 Hz, 1.2-0.63 A Adaptor DC output/LCD Monitor DC input 12.0 V dc, 4.0 A
Applicant Name and Address:	TVM CORP 7TH FL, 141 SEC 3 REN-AI RD TAIPEI 106 TAIWAN

This is to certify that representative samples of the products covered by this Test Report have been investigated in accordance with the above referenced Standards. The products have been found to comply with the requirements covering the category and the products are judged to be eligible for Follow-Up Service under the indicated Test Procedure. The manufacturer is authorized to use the UL Mark on such products which comply with this Test Report and any other applicable requirements of UL LLC ('UL') in accordance with the Follow-Up Service Agreement. Only those products which properly bear the UL Mark are considered as being covered by UL's Follow-Up Service under the indicated Test Procedure.

The applicant is authorized to reproduce the referenced Test Report provided it is reproduced in its entirety.

Any information and documentation involving UL Mark services are provided on behalf of UL LLC (UL) or any authorized licensee of UL.

Prepared by: David Shih

Reviewed by: Jimmy Deng

Supporting Documentation

The following documents located at the beginning of this Procedure supplement the requirements of this Test Report:

- A. Authorization - The Authorization page may include additional Factory Identification Code markings.
- B. Generic Inspection Instructions -
 - i. Part AC details important information which may be applicable to products covered by this Procedure. Products described in this Test Report must comply with any applicable items listed unless otherwise stated in the body of this Test Report.
 - ii. Part AE details any requirements which may be applicable to all products covered by this Procedure. Products described in this Test Report must comply with any applicable items listed unless otherwise stated in the body of each Test Report.
 - iii. Part AF details the requirements for the UL Certification Mark which is not controlled by the technical standard used to investigate these products. Products are permitted to bear only the Certification Mark(s) corresponding to the countries for which it is certified, as indicated in each Test Report.

Product Description

The 1928L is available in serial and USB(combo) touch interface as well as non-touch. The 1928L functionally consists of a 19" LCD main display with a touchscreen. The main display element is a 19" diagonal SXGA resolution (1280 x 1024) LCD display. The display consists of an LCD display and touchscreen. The 1928L is powered by 12 VDC from an external medical grade power source.

The 1928L is a medical display designed to present information to the operator and the customer.

Model Differences

N/A

Technical Considerations

- Classification of installation and use : Portable
- Device type (component/sub-assembly/ equipment/ system) : Equipment
- Intended use (Including type of patient, application location) : None
- Mode of operation : Continuous
- Supply connection : Appliance coupler
- Accessories and detachable parts included : None
- Other options include : None
- The product was investigated to the following additional standards:: CAN/CSA-C22.2 No. 60601-1 (2008) (Medical Electrical Equipment - Part 1: General Requirements for Basic Safety and Essential Performance) (includes National Differences for Canada), ANSI/AAMI ES60601-1 (2005 + C1:09 + A2:10) (Medical Electrical Equipment - Part 1: General Requirements for Basic Safety and Essential Performance) (includes Deviations for United States)
- The product was not investigated to the following standards or clauses:: Electromagnetic Compatibility (IEC 60601-1-2), Clause 14, Programmable Electronic Systems, Biocompatibility (ISO 10993-1)
- The degree of protection against harmful ingress of water is:: IPX1
- The mode of operation is:: Continuous
- The product is suitable for use in the presence of a flammable anesthetics mixture with air or oxygen or with nitrous oxide:: No

UL TEST REPORT AND PROCEDURE

Standard:	UL 60601-1, 1st Edition, 2006-04-26 (Medical Electrical Equipment, Part 1: General Requirements for Safety) CAN/CSA-C22.2 No. 601.1-M90, 2005 (Medical Electrical Equipment - Part 1: General Requirements for Safety)
Certification Type:	Classification
CCN:	PIDF, PIDF7 (Medical Electrical Equipment)
Product:	LCD Desktop Touchmonitor
Model:	ET1928L-XXXM-X-XXXXXXXXXX, where X may be A-Z, 0-9, hyphen or blank.
Rating:	LCD Desktop Touchmonitor: Input: 12 Vdc, 4.0 A Power Adaptor Hitron Electronics Corp. model:HES49-12040 Input:100-240 Vac, 1.0 A, 50/60 Hz Output: 12 V dc, 4 A. (Alternate) Power Adaptor Hitron Electronics Corp. model:HEMG49-S120400-7 Input: 100-240 V ac, 1.2-0.63A, 50/60 Hz Output: 12 V dc, 4 A.
Applicant Name and Address:	TVM CORP 7TH FL, 141 SEC 3 REN-AI RD TAIPEI 106 TAIWAN

This is to certify that representative samples of the products covered by this Test Report have been investigated in accordance with the above referenced Standards. The products have been found to comply with the requirements covering the category and the products are judged to be eligible for Follow-Up Service under the indicated Test Procedure. The manufacturer is authorized to use the UL Mark on such products which comply with this Test Report and any other applicable requirements of UL LLC ('UL') in accordance with the Follow-Up Service Agreement. Only those products which properly bear the UL Mark are considered as being covered by UL's Follow-Up Service under the indicated Test Procedure.

The applicant is authorized to reproduce the referenced Test Report provided it is reproduced in its entirety.

Any information and documentation involving UL Mark services are provided on behalf of UL LLC (UL) or any authorized licensee of UL.

Prepared by: David Shih

Reviewed by: Jimmy Deng

Copyright © 2012

Supporting Documentation

The following documents located at the beginning of this Procedure supplement the requirements of this Test Report:

- A. Authorization - The Authorization page may include additional Factory Identification Code markings.
- B. Generic Inspection Instructions -
 - i. Part AC details important information which may be applicable to products covered by this Procedure. Products described in this Test Report must comply with any applicable items listed unless otherwise stated in the body of this Test Report.
 - ii. Part AE details any requirements which may be applicable to all products covered by this Procedure. Products described in this Test Report must comply with any applicable items listed unless otherwise stated in the body of each Test Report.
 - iii. Part AF details the requirements for the UL Certification Mark which is not controlled by the technical standard used to investigate these products. Products are permitted to bear only the Certification Mark(s) corresponding to the countries for which it is certified, as indicated in each Test Report.

Product Description

The product covered is an accessory intended for use with Medical Equipment to display alpha, numerical and graphical data. The subject equipment consists of a LCD display comprising of amorphous silicon TFT attached to each signal electrode, a driving circuit, and a back light, powered by an external recognized AC/DC power adaptor (Hitron Electronics Corp., Model HES49-12040; QQHM2). (Aletnate: Hitron Electronics Corp. model:HEMG49-S120400-7)

Model Differences

N/A

Technical Considerations

- Classification of installation and use : Portable
- Supply connection : Appliance coupler
- Accessories and detachable parts included in the evaluation : None
- Options included : None
- The product was investigated to the following additional standards:: CAN/CSA-C22.2 No. 601.1-M90 (R2005) (includes National Differences for Canada), UL 60601-1, 1st Edition, 2006-04-26 (includes National Differences for USA),
- The product was not investigated to the following standards or clauses:: Clause 52.1, Programmable Electronic Systems (IEC 601-1-4) , Clause 48, Biocompatibility (ISO 10993-1) , Clause 36, Electromagnetic Compatibility (IEC 601-1-2)
- The product is Classified only to the following hazards:: Casualty , Shock , Fire
- The degree of protection against harmful ingress of water is:: Ordinary
- The mode of operation is:: Continuous
- Software is relied upon for meeting safety requirements related to mechanical, fire and shock:: No
- The product is suitable for use in the presence of a flammable anesthetics mixture with air or oxygen or with nitrous oxide:: No

Additional Information

Project 06CA59197, E230760-A2, Amendment 1:
1. To alternate a Humidity Preconditioning Treatment Test in IPX1 test condition.

Project 07CB07338, E230760-A2, Amendment 2:

1. To modify the information of IPX1 in Label and User Manual.

Project 07CA14754, E230760-A2, Amendment 3:

1. To add VESA Mounting Function and Change Base to be optional.

Project 08CA03217, E230760-A2-UL-2, Reissue 1:

1. Change front enclosure to "MIN HB".

Project 10CA04218, E230760-A2-UL-2, Reissue 1_Amendment 1:

Alternate a LCD panel, Samsung/model LTM190E4-L02

Test Record 3 (12CA25519)

Alternate a Power Adaptor, Hitron Electronics Corp. model:HEMG49-S120400-7

Additional Standards

The product fulfills the requirements of: CAN/CSA-C22.2 No. 601.1-M90 (R2005) (includes National Differences for Canada)

Markings and instructions

Clause Title	Marking or Instruction Details
Company identification	Classified or Recognized company's name, Trade name, Trademark or File
Model	Model number
Supply Connection	Voltage range, ac/dc, phases if more than single phase
Supply Frequency	Rated frequency range in hertz
Power Input	Amps, VA, or Watts
IP Rating	IPX1
Mode of Operation	Continuous
Attention, consult accompanying documents	
US Hospital Grade Marking	"Grounding Reliability Can Only Be Achieved When The Equipment Is Connected To An Equivalent Receptacle Marked 'Hospital Only' Or 'Hospital Grade'." (located on product or power supply cord)

Special Instructions to UL Representative

N/A



中国国家强制性产品认证证书

证书编号: 2010010903427943

委托人名称、地址

Elo TouchSystems, Inc
301 Constitution Drive, Menlo Park, CA 94025-1111 USA

生产者(制造商)名称、地址

Elo TouchSystems, Inc
301 Constitution Drive, Menlo Park, CA 94025-1111 USA

生产企业名称、地址

钛积光电(厦门)有限公司
厦门市集美北部工业区天阳路1-7号

产品名称和系列、规格、型号

液晶触控显示器

ET1928L-***M-*-*-*-*-*-*-*-* (*~0~9, A~Z, 连字符或空白, *代表销售区域不同, 不影响安全和电磁兼容) 12VDC, 4A Class III (电源适配器(可选): HES49-12040 输出: 12VDC 4A; HEMG49-S120400-7 输出: 12VDC, 4A) (电线组件可选)

产品标准和技术要求

GB9254-2008 GB4943-2001 GB17625.1-2003

上述产品符合强制性产品认证实施规则
CNCA-01C-020: 2010的要求, 特发此证。

发证日期: 2012年06月08日

有效期至: 2015年01月28日

证书有效期内本证书的有效性依据发证机构的定期监督获得保持。

本证书为变更证书, 证书首次颁发日期: 2010年08月30日

本证书的相关信息可通过国家认监委网站www.cnca.gov.cn查询



主任:

中国质量认证中心

中国·北京·南四环西路188号9区 100070

<http://www.cqc.com.cn>



Q 0546226



CERTIFICATE FOR CHINA COMPULSORY PRODUCT CERTIFICATION

No. : 2010010903427943

NAME AND ADDRESS OF THE APPLICANT

Elo TouchSystems, Inc
301 Constitution Drive, Menlo Park, CA 94025-1111 USA

NAME AND ADDRESS OF THE MANUFACTURER

Elo TouchSystems, Inc
301 Constitution Drive, Menlo Park, CA 94025-1111 USA

NAME AND ADDRESS OF THE FACTORY

Touch Video Monitor Corp.
No 1-7 Tian Yang Rd Jimei N Industrial Park Xiamen P. R. China

NAME, MODEL AND SPECIFICATION

LCD Touch Monitor

ET1928L-***M-**-***** (*=0~9, A~Z, 连字符或空白, *代表销售区域不同, 不影响安全和电磁兼容) 12VDC, 4A Class III (电源适配器(可选): HES49-12040 输出: 12VDC 4A;HEMG49-S120400-7 输出: 12VDC, 4A) (电线组件可选)

THE STANDARDS AND TECHNICAL REQUIREMENTS FOR THE PRODUCTS

GB9254-2008 GB4943-2001 GB17625.1-2003

This is to certify that the above mentioned products have met the requirements of implementation rules for compulsory certification(REF NO. CNCA-01C-020:2010).

Date of issue: Jun. 08, 2012 Date of expiry: Jan. 28, 2015

Validity of this certificate is subject to positive result of the regular follow up inspection by issuing certification body until the expiry date.

Date of original certification: Aug. 30, 2010

This certificate can be verified through CNCA's website: www.cnca.gov.cn



President:

Wang Kejiao

CHINA QUALITY CERTIFICATION CENTRE

Section 9, No. 188, Nansihuan Xilu, Beijing 100070 P.R. China

<http://www.cqc.com.cn>



Q 0546226



中国国家强制性产品认证证书

变更结论

Modification Approval

证书编号Certificate No.: 2010010903427943

申请编号Application No.: A2012CCC0903-1288733

申请人Applicant: Elo TouchSystems, Inc Elo TouchSystems, Inc

证书编号Certificate No. (变更前Original): 2010010903427943

证书编号Certificate No. (变更后Present): 2010010903427943

变更产品名称 液晶触控显示器

Name of Product Modified: LCD Touch Monitor

型号Model/Type: ET1928L-***M-***** (*=0~9, A~Z, 连字符或空白, *代表销售区域不同, 不影响安全和电磁兼容) 12VDC, 4A Class III (电源适配器 (可选): HES49-12040 输出: 12VDC 4A; HEMG49-S120400-7 输出: 12VDC, 4A) (电线组件可选)

变更情况Situation of Modification:

见附件See Annex:

序号	项目Item	变更前Original	变更后Present
1	3. 删除原报告中适配器, 新增一款适配器, 适配器并变更为可选	见原报告 C-003-20101630 (A2010CC C0903-962641)	电源适配器 (可选) 1. 型号: HES49-12040 输入: 100-240 VAC, 50/60 Hz, 1.0 A, 输出: 12VDC, 4A (产品销售时不配备电线组 件) 生产者: 海昌电子股份有限公司 生产企业: 福州山昌电子有限公司 认证标准: GB4943-2001, GB9254-2008, GB17625.1-2003 CCC证书号: 2007010907241678

第 1 页 / 共 4 页



主 任:

中国质量认证中心

中国 · 北京 · 南四环西路 188 号 9 区 100070

<http://www.cqc.com.cn>





中国国家强制性产品认证证书

变更结论

Modification Approval

证书编号Certificate No.: 2010010903427943

			2. 型号: HEMG49-S120400-7 输入: 100-240Vac, 60/50Hz, 1.2-0.63A 输出: 12Vdc, 4A, 48W max (产品不带电线组件销售) 生产者: 海昌电子股份有限公司 生产企业: 福州山昌电子有限公司 认证标准: GB4943-2001, GB9254-2008, GB17625.1-2003 CCC证书号: 2012010907543076
2	4. 修正EMC清单中信号线中的磁环 (VGA) (可选)	见原报告 C-003-20101630(A2010CC C0903-962641)	1. 型号: F2 规格: 14.2*28.5*7.2 mm 制造商: 湖州科峰磁业有限公司 2. 型号: F12 规格: 12*20*6.5 mm 制造商: 湖州科峰磁业有限公司 3. 型号: F5B 规格: 14.2*28.5*6.7 mm 制造商: 优磁电子有限公司 4. 型号: YM 规格: 14.2*28.5*7.2 mm 制造商: 苏州阳山电子有限公司
3	5. 修正EMC清单中信号线中的磁环 (DVI) (可选)	见原报告 C-003-20101630(A2010CC C0903-962641)	1. 型号: F2 规格: 14.2*28.5*7.2 mm 制造商: 湖州科峰磁业有限公司 2. 型号: F12 规格: 12*20*6.5 mm 制造商: 湖州科峰磁业有限公司

第 2 页 / 共 4 页



主任:

中国质量认证中心

中国 · 北京 · 南四环西路 188 号 9 区 100070

<http://www.cqc.com.cn>





中国国家强制性产品认证证书

变更结论

Modification Approval

证书编号Certificate No.: 2010010903427943

			3. 型号: F5B 规格: 14.2*28.5*6.7 mm 制造商: 优磁电子有限公司 4. 型号: YM 规格: 14.2*28.5*7.2 mm 制造商: 苏州阳山电子有限公司
4	6. 在EMC清单中补充显示屏 (不含LCD控制电路) 信息	见原报告 C-003-20101630 (A2010CC C0903-962641)	1. 型号: M190EG01 规格: 19" TFT-LCD, V-1min. 制造商: AU OPTRONICS CORPORATION 2. 型号: LTM190E4-L02 规格: 19" TFT-LCD, V-1min. 制造商: Samsung
5	7. 修正LCD控制电路信息	见原报告 C-003-20101630 (A2010CC C0903-962641)	集成在主板上
6	8. 增加一款铭牌样式	见原报告 C-003-20101630 (A2010CC C0903-962641)	见申请编号: A2012CCC0903-1288733检测机 构变更确认表
7	1. 实施规则版本变更	CNCA-01C-020:2007	CNCA-01C-020:2010
8	2. 增加电线组件, 并为修 正为 (可选)	见原报告 C-003-20101630 (A20	电线组件 (可选) 1. 型号: LSG-31+LS-70, LSG-31L+L

第 3 页 / 共 4 页



主 任:

中国质量认证中心

中国 · 北京 · 南四环西路 188 号 9 区 100070

<http://www.cqc.com.cn>





中国国家强制性产品认证证书

变更结论

Modification Approval

证书编号Certificate No.: 2010010903427943

		10CCC0903-962641)	S-70L, LSG-31L+LS-70, LSG-31+LS-70L 规格: 250V, 10A(配60227IEC53, 3X0.75平方毫米, 3X1平方毫米) 生产者: 宝安松岗良维电子厂 生产企业: 宝安松岗良维电子厂 认证标准: GB15934-2008 CCC证书号: 2009010101348285 本次增加: 2. 型号:SP-504+IS-14, SP-504+IS-15, SP-506+IS-14, SP-506+IS-15, SP-506A+IS-14, SP-506A+IS-15 规格:250V, 10A(配60227IEC53 3X0.75平方毫米, 3X1平方毫米) 生产者: 深圳市宝安区松岗镒胜制品厂 生产企业: 深圳市宝安区松岗镒胜制品厂 认证标准: GB15934-2008 CCC证书号: 2007010101232000
--	--	-------------------	--

Above Modification is approved. The present will replace the original.

现批准以上变更。

批准日期 (Date): 2012-06-08

第 4 页 / 共 4 页



主任:

中国质量认证中心

中国 · 北京 · 南四环西路 188 号 9 区 100070

<http://www.cqc.com.cn>



CERTIFICATE OF CONFORMITY




Equipment : 19" LCD Touch monitor
Brand Name : TE Connectivity, Elo, TPS, TYCO
Test Model No. : ET1928L-8CJM-1-BG-NKD-G
Multiple Listing : ET1928LXXXXXXXXXXXXXXXXXXXXX
(X= 0-9, A-Z, hyphen or blank)
Applicant : Elo TouchSystems Inc.
Test Report No. : CE970618A12D

We, **Bureau Veritas Consumer Products Services (H.K.) Ltd., Taoyuan Branch**, declare that the equipment above has been tested in our facility and found compliance with the requirement limits of applicable standards. The test record, data evaluation and Equipment Under Test (EUT) configurations represented herein are true and accurate under the standards herein specified.

CISPR 22:2008, Class B	EN 61000-6-1: 2007
EN 55022:2010, Class B	CISPR 24:2010
AS/NZS CISPR 22:2009, Class B	EN 55024:2010
EN 61000-3-2:2006 +A1:2009 +A2:2009	IEC 61000-4-2:2008 ED.2.0
EN 61000-3-3:2008	IEC 61000-4-3:2010 ED.3.2
	IEC 61000-4-4:2012 ED.3.0
	IEC 61000-4-5:2005 ED.2.0
	IEC 61000-4-6:2008 ED.3.0
	IEC 61000-4-8:2009 ED.2.0
	IEC 61000-4-11:2004 ED.2.0
	ENV 50204: 1995

Notes: The above IEC basic standards are applied with latest version if customer has no special requirement
In accordance with the council directive 2004/108/EC


Kenny Meng / Assistant Manager
Sep. 21, 2012



No. 47, 14th Ling, Chia Pau Vil., Lin Kou Dist., New Taipei City, Taiwan (R.O.C.)
Tel: 886-2-26052180 Fax: 886-2-26051924
<http://www.adt.com.tw> E-Mail: service.adt@tw.bureauveritas.com

CERTIFICATE OF CONFORMITY



A D T



Equipment : 19" LCD Touch monitor
Brand Name : TE Connectivity, Elo, TPS, TYCO
Test Model No. : ET1928L-8CJM-1-BG-NKD-G
Multiple Listing : ET1928LXXXXXXXXXXXXXXXXXXXXX
(X= 0-9, A-Z, hyphen or blank)
Applicant : Elo TouchSystems Inc.
Test Report No. : CE970618A12C

We, **Bureau Veritas Consumer Products Services (H.K.) Ltd., Taoyuan Branch**, declare that the equipment above has been tested in our facility and found compliance with the requirement limits of applicable standards. The test record, data evaluation and Equipment Under Test (EUT) configurations represented herein are true and accurate under the standards herein specified.

EN 60601-1-2:2007

The EMC limits and test methods are referred to the following standards:

Emission:

Immunity:

CISPR 11:2009 +A1:2010 (Group I, Class B) IEC 61000-4-2:2008 ED.2.0

AS/NZS CISPR 11:2011 (Group I, Class B) IEC 61000-4-3:2010 ED.3.2

EN 55011:2009 +A1:2010, (Group I, Class B) IEC 61000-4-4:2011 ED.2.1

IEC 61000-3-2:2005 +A1:2008 +A2:2009 IEC 61000-4-5:2005 ED.2.0


IEC 61000-3-3:2008 IEC 61000-4-6:2008 ED.3.0

IEC 61000-4-8:2009 ED.2.0

IEC 61000-4-11:2004 ED.2.0

ENV 50204:1995

Notes: The above IEC basic standards are applied with latest version if customer has no special requirement
In accordance with the council directive 2004/108/EC


Kenny Meng / Assistant Manager
Mar. 27, 2012



No. 47, 14th Ling, Chia Pau Vil., Lin Kou Dist., New Taipei City, Taiwan (R.O.C.)

Tel: 886-2-26052180 Fax: 886-2-26051924

<http://www.adt.com.tw> E-Mail: service.adt@tw.bureauveritas.com

CERTIFICATE OF CONFORMITY



A D T


Equipment : 19" LCD Touch monitor
Brand Name : TE Connectivity, Elo, TPS, TYCO
Test Model No. : ET1928L-8CJM-1-BG-NKD-G
Multiple Listing : ET1928LXXXXXXXXXXXXXXXXXXXXX
(X= 0-9, A-Z, hyphen or blank)
Applicant : Elo TouchSystems Inc.
Test Report No. : FD970618A12C

We, **Bureau Veritas Consumer Products Services (H.K.) Ltd., Taoyuan Branch**, declare that the equipment above has been tested in our facility and found compliance with the requirement limits of applicable standards. The test record, data evaluation and Equipment Under Test (EUT) configurations represented herein are true and accurate under the standards herein specified.

FCC Part 15, Subpart B, Class B

ICES-003: 2004, Class B

ANSI C63.4-2003


Kenny Meng / Assistant Manager
Mar. 27, 2012



No. 47, 14th Ling, Chia Pau Vil., Lin Kou Dist., New Taipei City, Taiwan (R.O.C.)

Tel: 886-2-26052180 Fax: 886-2-26051924

<http://www.adt.com.tw> E-Mail: service.adt@tw.bureauveritas.com

변경신고 처리결과 통보서

전자민원신청번호		201215701000052350	접수일자	2012-04-04
대상 기기	상호 또는 성명	TYCO ELECTRONICS, ELO TOUCHSYSTEMS	적합성평가 분야	적합인증
	기자재의 명칭	LCD Touch Monitor	기본모델명	ET1928L-7CAM-1-GY-G
	적합성평가 번호	TY1-ET1928L7CAM1GY	적합성평가 연 월 일	2012-04-09
변경 사항		변경 전		변경 후
기술기준		Adaptor:HES49-12040 / Taiwan HiTRON electronics corporation		Adaptor:HEMG49-S120400-7 / China HiTRON electronics corporation

「방송통신기자재등의 적합성평가에 관한 고시」 제16조에 따른 적합성평가
사항의 변경신고 전에 대하여 위와 같이 변경처리 되었음을 알려드립니다.

2012년(Year) 04월(Month) 09일(Date)

국립전파연구원장





經濟部標準檢驗局

BUREAU OF STANDARDS METROLOGY AND INSPECTION
MINISTRY OF ECONOMIC AFFAIRS



商品驗證登錄證書

CERTIFICATE OF THE REGISTRATION OF PRODUCT CERTIFICATION

證書號碼： CI311065240622 號 00

Certificate No.

茲據

台灣泰科電子股份有限公司

申請驗證登錄，經審查結果符合規

定，准予登錄並使用商品安全標章及識別號碼：

R31524

。其登錄事項如下：

The application made by

for Registration of Product

Certification has been reviewed and found to be in compliance with related regulations. Therefore, registration is granted with the

Product Safety Mark and the Identification No.

R31524.

Details of the registration are follows:

申請人：台灣泰科電子股份有限公司

統一編號：22522138

Applicant

Uniform No.

地址：臺北市信義區東興路45號3樓

Address

生產廠場：鈦積光電(廈門)有限公司

Factory

廠址：福建省廈門市集美北區工業區天陽路1-7號

Factory Address

商品種類名稱：

Type/name of product

商品分類號列：8528.59.10.00.5

C. C. C Code

中文名稱：液晶觸控顯示器

Chinese name

英文名稱：LCD Touch Monitor

English name

型式：ET1928L-8CWM

Type

系列型式：詳如附表

Series of the type

依據標準：CNS14336-1 (99年版)、CNS13438 (95年完整版)

Standards

標準檢驗局或所屬分局發證

(本證書經驗證機關使用鋼印後生效)

This certificate shall be issued by the BSMI or its branches.

(This certificate will become effective only when stamped with the BSMI seal.)

登錄日期：中華民國	101	年	06	月	18	日
Registration Date	2012	(year)	06	(month)	18	(day)
本證書有效期限至	104	年	06	月	17	日
Expiration Date	2015	(year)	06	(month)	17	(day)
發證日期：中華民國	101	年	06	月	18	日
Date of issue	2012	(year)	06	(month)	18	(day)

註1：持本證書進口驗證登錄商品時，進口人須與本證書名義人相同。

註2：次年度商品驗證登錄年費繳納期限為當年11月30日，逾期未繳納者，經限期繳納屆期未繳納，即依商品檢驗法第42條第7款規定廢止驗證登錄，並自次年度1月1日起生效。

經濟部標準檢驗局
BUREAU OF STANDARDS METROLOGY AND INSPECTION
MINISTRY OF ECONOMIC AFFAIRS

商品驗證登錄證書

CERTIFICATE OF THE REGISTRATION OF PRODUCT CERTIFICATION

證書號碼： CI311065240622 號 00

Certificate No.

系列型式： ET1928L-ONAM-1-BG-3-G

Series of the type

ET1928L-ONAM-1-GY-3-G

ET1928L-ONWM-1-BG-3-G

ET1928L-ONWM-1-GY-3-G

ET1928L-ONWM-1-ROGH2-3-G

ET1928L-ONWM-1-ROGH3-3-G

ET1928L-7CAM-1-BG-3-G

ET1928L-7CAM-1-GY-3-G

ET1928L-7CEM-1-ROAF-3-G

ET1928L-7CJM-1-BG-3-G

ET1928L-7CWM

ET1928L-7CWM-1-BG-3-G

ET1928L-7CWM-1-GY-3-G

ET1928L-7CWM-1-RROB-G

ET1928L-8CAM-1-BG-3-G

ET1928L-8CAM-1-GY-3-G

ET1928L-8CEM-1-ROGG-3-G

ET1928L-8CJA-1-BG-3-G

ET1928L-8CJA-1-GY-3-G

ET1928L-8CJM-1-BG-3-G

ET1928L-8CJM-1-BG-NKD-G

ET1928L-8CJM-1-GY-3-G

ET1928L-8CJN-1-BG-NKD-G

ET1928L-8CWM

ET1928L-8CWM-1-BG-3-G

ET1928L-8CWM-1-GY-3-G

ET1928L-8CWM-1-RDAT-3-G

ET1928L-8CWM-1-RMHJ-G

ET1928L-8CWM-1-ROGH-3-G

ET1928L-8CWM-1-ROGH2-3-G

ET1928L-8CWM-1-ROGH3-3-G

ET1928L-8CWM-1-RSTU-G

ET1928L-AUAM-1-BG-3-G

ET1928L-AUAM-1-GY-3-G

ET1928L-AUWM

ET1928L-AUWM-1-BG-3-G

ET1928L-AUWM-1-GY-3-G

(以下空白)

1 CERTIFICATION

PRODUCT: 19" LCD Touch monitor
BRAND NAME: TE Connectivity, Elo, TPS, TYCO
MODEL NO.: ET1928LXXXXXXXXXXXXXXXXXXXXX
(X= 0-9, A-Z, hyphen or blank)
TEST ITEM: ENGINEERING SAMPLE
APPLICANT: Touch Panel Systems K.K.
TESTED: Mar. 13 ~ 15, 2012
STANDARDS: V-3/2011.04, TECHNICAL REQUIREMENTS, Class B
V-4/2010.04, SUPPLEMENTARY TEST CONDITIONS
FOR EQUIPMENT UNDER TEST

The above equipment (Model No.: ET1928L-8CJM-1-BG-NKD-G) has been tested by **Bureau Veritas Consumer Products Services (H.K.) Ltd., Taoyuan Branch**, and found compliance with the requirement of the above standards. The test record, data evaluation & Equipment Under Test (EUT) configurations represented herein are true and accurate accounts of the measurements of the sample's EMC characteristics under the conditions specified in this report.

PREPARED BY : Albee Chu, **DATE:** Mar. 26, 2012
(Albee Chu / Specialist)

APPROVED BY : Kenny Meng, **DATE:** Mar. 26, 2012
(Kenny Meng / Assistant Manager)

2 SUMMARY OF TEST RESULTS

The EUT has been tested according to the following specifications:

Standard	Test Type	Result	Remarks
V-3/2011.04, Class B	Conducted Test	PASS	Meets Class B Limit Minimum passing margin is –14.14 dB at 1.14063 MHz
	Conducted Test (telecom port)	N/A	Not Applicable
	Radiated Test (30MHz ~ 2GHz)	PASS	Meets Class B Limit Minimum passing margin is –3.95 dB at 131.34 MHz

Note: The EUT highest frequency generated **135MHz** and therefore the test frequency range was performed up to 2GHz for radiated emission test.

2.1 MEASUREMENT UNCERTAINTY

Where relevant, the following measurement uncertainty levels have been estimated for tests performed on the EUT as specified in CISPR 16-4-2:

This uncertainty represents an expanded uncertainty expressed at approximately the 95% confidence level using a coverage factor of k=2.

MEASUREMENT	FREQUENCY	UNCERTAINTY	VCCI Site Registration No.
Conducted emissions	150kHz ~ 30MHz	3.43 dB	C-1852
Radiated emissions	30MHz ~ 1GHz	3.24 dB	R-1248
	Above 1GHz	3.36 dB	G-257